

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 000067  
9:19 17/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:14  
PRESSURIZING 3:16  
STERILIZING 4:40  
136.4 C 224kPa 4:40  
136.4 C 225kPa 5:10  
136.3 C 223kPa 5:40  
136.2 C 221kPa 6:10  
136.4 C 224kPa 6:40  
136.4 C 225kPa 7:10  
136.3 C 222kPa 7:40  
136.1 C 221kPa 8:10  
136.1 C 221kPa 8:10  
VENTING 8:10  
AIR DRYING 9:01  
CYCLE COMPLETE 19:57  
-----

Digital Signature #  
A1AFD37A5D900EBA  
-----