
Midmark M11 - v1.0.0
Total Cycles: 2060

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.2 F 0.0
2:00 174.3 F 3.5
4:00 185.4 F 7.6
6:00 207.3 F 12.6
8:00 219.6 F 7.0
10:00 233.2 F 12.2
12:00 234.5 F 9.1
14:00 237.9 F 9.5
16:00 246.8 F 13.9
18:00 259.3 F 20.7
20:00 270.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 272.1 F 28.9
3:00 271.7 F 28.9
4:00 272.7 F 29.4

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 209.8 F 0.0
10:00 227.4 F 0.0
15:00 228.7 F 0.0
20:00 227.2 F 0.0
25:00 223.5 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 20:54
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

06/27/2022 12:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____