
Midmark M11 - v1.0.0
Total Cycles: 1045

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2021 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 183.7 F 3.6
4:00 204.0 F 8.0
6:00 222.8 F 13.5
8:00 230.9 F 8.8
10:00 240.4 F 10.6
12:00 242.5 F 12.0
14:00 240.9 F 10.7
16:00 254.4 F 17.5
18:00 265.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.8
2:00 271.5 F 28.2
3:00 272.5 F 28.8
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 218.5 F 0.0
10:00 218.8 F 0.0
15:00 212.1 F 0.0
20:00 207.9 F 0.0
25:00 206.8 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 2:45
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

07/22/2021 11:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____