
Midmark M11 - v1.0.0
Total Cycles: 5372

Name: RITTER 1
Sterilizer ID: V2178198
02/26/2025 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 197.3 F 4.4
4:00 208.8 F 10.4
6:00 220.5 F 5.4
8:00 238.8 F 11.3
10:00 236.1 F 9.0
12:00 238.2 F 9.1
14:00 249.9 F 15.1
16:00 263.0 F 22.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 271.9 F 28.2
2:00 272.1 F 28.4
3:00 272.4 F 28.7
4:00 272.6 F 29.0

Min 271.7 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 223.5 F 0.0
10:00 217.5 F 0.0
15:00 226.4 F 0.0
20:00 221.6 F 0.0
25:00 222.3 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 3:20
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

02/26/2025 06:08 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____