
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005672
14:32 05/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 4:03
136.4 C 220kPa 4:03
136.4 C 225kPa 4:33
136.2 C 224kPa 5:03
136.3 C 224kPa 5:33
136.0 C 223kPa 6:03
136.4 C 225kPa 6:33
136.1 C 222kPa 7:03
136.4 C 225kPa 7:33
136.4 C 225kPa 7:33
VENTING 7:33
AIR DRYING 8:35
CYCLE COMPLETE 8:41

Digital Signature #
7E540454653D9BB1
