
Midmark M11 - v1.0.0
Total Cycles: 5212

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2024 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.4 F 0.0
2:00 126.4 F 0.0
4:00 151.6 F 0.0
6:00 182.0 F 0.0
8:00 201.1 F 0.2
10:00 212.7 F 0.7
12:00 223.3 F 4.4
14:00 231.1 F 7.6
16:00 239.8 F 11.3
18:00 234.7 F 8.1
20:00 246.9 F 13.1
22:00 244.8 F 12.3
24:00 242.4 F 11.2
26:00 254.0 F 17.0
28:00 263.7 F 22.7
30:00 272.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 271.8 F 28.0
2:00 271.9 F 28.1
3:00 272.1 F 28.3
4:00 272.3 F 28.4

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 206.4 F 0.0
10:00 205.3 F 0.0
15:00 205.8 F 0.0
20:00 204.3 F 0.0
25:00 204.7 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

12/20/2024 04:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____