
Midmark M11 - v1.0.0
Total Cycles: 3020

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2023 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.4 F 0.0
2:00 142.9 F 0.5
4:00 170.6 F 2.4
6:00 197.3 F 5.6
8:00 211.4 F 10.1
10:00 223.8 F 6.0
12:00 230.9 F 8.9
14:00 241.1 F 12.3
16:00 238.5 F 10.6
18:00 235.9 F 8.2
20:00 249.7 F 15.1
22:00 261.5 F 21.8
24:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.6 F 28.6
3:00 272.4 F 28.9
4:00 271.5 F 28.3

Min 271.4 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 220.7 F 0.0
10:00 233.5 F 0.0
15:00 234.7 F 0.0
20:00 232.6 F 0.0
25:00 229.1 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:36
CYCLE COMPLETE

03/20/2023 02:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____