
Midmark M11 - v1.0.0
Total Cycles: 671

Name: MIDMARK
Sterilizer ID: V2178198
02/17/2021 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.5 F 0.0
2:00 168.9 F 3.3
4:00 182.6 F 8.0
6:00 211.2 F 10.3
8:00 223.0 F 9.7
10:00 232.4 F 7.6
12:00 245.1 F 13.5
14:00 244.0 F 12.1
16:00 257.7 F 19.3
18:00 269.6 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.1 F 29.0
2:00 272.5 F 29.2
3:00 272.6 F 29.2
4:00 272.5 F 29.2

Min 271.8 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 231.7 F 0.0
10:00 229.0 F 0.0
15:00 227.2 F 0.0
20:00 227.2 F 0.0
Dry Cycle Aborted

FILLING: 1:13
HEATING: 19:27
STERILIZING: 4:00
VENTING: 2:56
DRYING: 21:31
TOTAL CYCLE: 00:49:07

02/17/2021 11:56 AM
ACCEPTED: X
REJECTED:
Operator: U01