

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3231

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2023 08:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 70.3 F 0.0  
4:00 78.0 F 0.2  
6:00 102.0 F 1.2  
8:00 146.0 F 3.9  
10:00 170.6 F 6.9  
12:00 191.9 F 10.2  
14:00 223.4 F 6.2  
16:00 227.4 F 8.1  
18:00 235.1 F 12.2  
20:00 232.6 F 8.2  
22:00 243.8 F 13.4  
24:00 241.6 F 11.5  
26:00 253.6 F 17.7  
28:00 264.4 F 24.1  
30:00 270.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.1 F 29.4  
2:00 272.4 F 29.7  
3:00 272.5 F 29.9  
4:00 272.3 F 30.1  
  
Min 271.4 F 29.3  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.9 F 0.0  
5:00 237.7 F 0.0  
10:00 208.3 F 0.0  
15:00 195.7 F 0.0  
20:00 172.2 F 0.0  
25:00 171.0 F 0.0  
30:00 170.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:44  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:08:49  
CYCLE COMPLETE

12/19/2023 09:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_