
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001213
8:30 19/02/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:52
STERILIZING 4:56
136.4 C 223kPa 4:56
136.3 C 224kPa 5:26
135.9 C 220kPa 5:56
136.4 C 224kPa 6:26
136.4 C 223kPa 6:56
136.2 C 223kPa 7:26
136.4 C 223kPa 7:56
136.5 C 223kPa 8:26
136.5 C 223kPa 8:26
VENTING 8:26
AIR DRYING 9:21
CYCLE COMPLETE 9:41

Digital Signature #
547610FDF57C8704
