
Midmark M11 - v1.0.0
Total Cycles: 3722

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2023 07:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 71.8 F 0.0
4:00 75.8 F 0.0
6:00 102.0 F 0.0
8:00 142.8 F 1.6
10:00 172.8 F 3.5
12:00 188.8 F 6.6
14:00 205.2 F 10.1
16:00 215.7 F 9.4
18:00 226.5 F 7.4
20:00 237.0 F 11.3
22:00 230.4 F 7.1
24:00 244.4 F 12.7
26:00 241.2 F 10.8
28:00 253.8 F 17.2
30:00 264.9 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 28.9
2:00 271.5 F 28.1
3:00 272.5 F 29.0
4:00 272.1 F 28.5

Min 271.1 F 28.0
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 230.1 F 0.0
10:00 208.2 F 0.0
15:00 232.2 F 0.0
20:00 224.5 F 0.0
25:00 209.0 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:57
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

09/06/2023 08:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____