
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001538
13:17 10/28/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:34
PRESSURIZING 6:16
STERILIZING 7:48
136.0 C 220kPa 7:48
136.4 C 223kPa 8:18
136.5 C 219kPa 8:48
136.1 C 219kPa 9:18
136.4 C 223kPa 9:48
136.4 C 222kPa 10:18
136.4 C 220kPa 10:48
136.4 C 222kPa 11:18
VENTING 11:18
AIR DRYING 12:18
CYCLE COMPLETE 42:18

Digital Signature #
58DD5F9D11C9944F
