

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2766

Name: RITTER 1  
Sterilizer ID: V2178198  
01/11/2023 12:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
2:00 200.1 F 0.0  
4:00 215.3 F 5.5  
6:00 229.8 F 10.9  
8:00 231.8 F 6.8  
10:00 243.0 F 12.8  
12:00 241.4 F 10.9  
14:00 239.4 F 9.7  
16:00 253.9 F 16.8  
18:00 265.7 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 27.8  
1:00 272.0 F 28.5  
2:00 272.1 F 28.2  
3:00 271.4 F 27.9  
4:00 272.5 F 28.6  
  
Min 271.3 F 27.7  
Max 272.5 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 234.4 F 0.0  
10:00 230.9 F 0.0  
15:00 226.4 F 0.0  
20:00 228.2 F 0.0  
25:00 230.8 F 0.0  
30:00 224.7 F 0.0  
  
Drying Complete

FILLING: 1:34  
HEATING: 19:35  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

01/11/2023 01:48 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_