

-----  
Midmark M11 - v1.0.5  
Total Cycles: 898

Name: RITTER2  
Sterilizer ID: V2313228  
03/24/2022 06:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.2 F 0.0  
2:00 86.4 F 0.0  
4:00 81.4 F 0.0  
6:00 101.8 F 0.0  
8:00 159.5 F 1.1  
10:00 189.0 F 1.9  
12:00 209.1 F 3.1  
14:00 218.7 F 6.5  
16:00 228.3 F 10.1  
18:00 234.2 F 9.8  
20:00 238.2 F 10.6  
22:00 235.9 F 8.4  
24:00 242.6 F 11.2  
26:00 246.8 F 13.9  
28:00 259.3 F 20.7  
30:00 270.2 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.4 F 29.0  
2:00 272.3 F 29.3  
3:00 272.0 F 28.7  
4:00 272.7 F 29.4

Min 271.6 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
5:00 207.7 F 0.0  
10:00 207.8 F 0.0  
15:00 202.1 F 0.0  
20:00 187.0 F 0.0  
25:00 190.9 F 0.0  
30:00 187.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:51  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:50  
CYCLE COMPLETE

03/24/2022 08:33 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02