
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001602
12:18 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:12
STERILIZING 3:09
136.0 C 220kPa 3:09
136.4 C 226kPa 3:39
136.2 C 223kPa 4:09
136.4 C 223kPa 4:39
136.4 C 225kPa 5:09
136.4 C 226kPa 5:39
136.2 C 224kPa 6:09
136.4 C 226kPa 6:39
136.4 C 226kPa 6:39
VENTING 6:39
AIR DRYING 7:33
CYCLE COMPLETE 7:45

Digital Signature #
303B5BBBF2EE49C7
