
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007112
10:39 26/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:29
STERILIZING 3:26
136.0 C 220kPa 3:26
136.4 C 225kPa 3:56
136.1 C 222kPa 4:26
136.2 C 223kPa 4:56
136.4 C 225kPa 5:26
136.4 C 223kPa 5:56
136.4 C 221kPa 6:26
136.3 C 222kPa 6:56
136.3 C 222kPa 6:56
VENTING 6:57
AIR DRYING 7:52
CYCLE COMPLETE 7:56

Digital Signature #
B9D88932C66EAE5B
