
Midmark M11 - v1.0.5
Total Cycles: 3074

Name: RITTER2
Sterilizer ID: V2313228
11/01/2023 05:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.6 F 0.0
2:00 194.3 F 5.0
4:00 219.3 F 12.3
6:00 232.2 F 9.1
8:00 236.1 F 8.6
10:00 239.8 F 9.9
12:00 251.9 F 16.4
14:00 266.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.1 F 28.6
2:00 272.7 F 29.0
3:00 272.4 F 29.3
4:00 272.1 F 28.6

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 212.4 F 0.0
10:00 217.2 F 0.0
15:00 220.1 F 0.0
20:00 216.7 F 0.0
25:00 214.1 F 0.0
30:00 212.9 F 0.0

Drying Complete

FILLING: 2:47
HEATING: 15:17
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:55:07
CYCLE COMPLETE

11/01/2023 06:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____