
Midmark M11 - v1.0.5
Total Cycles: 2815

Name: RITTER2
Sterilizer ID: V2313228
08/30/2023 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 78.8 F 0.0
4:00 83.1 F 0.0
6:00 105.3 F 0.0
8:00 143.4 F 0.0
10:00 196.2 F 0.0
12:00 202.5 F 0.9
14:00 213.3 F 3.7
16:00 221.7 F 6.5
18:00 230.7 F 9.8
20:00 240.3 F 12.3
22:00 237.9 F 10.1
24:00 237.7 F 9.5
26:00 245.8 F 13.5
28:00 244.3 F 12.6
30:00 256.1 F 18.8
32:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.2
2:00 271.7 F 28.9
3:00 272.4 F 29.1
4:00 271.6 F 28.9

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 224.8 F 0.0
10:00 228.8 F 0.0
15:00 218.5 F 0.0
20:00 192.7 F 0.0
25:00 231.2 F 0.0
30:00 227.9 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:39
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:11:52
CYCLE COMPLETE

08/30/2023 09:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____