

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3304

Name: RITTER2  
Sterilizer ID: V2313228  
01/16/2024 02:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.8 F 0.0  
2:00 85.0 F 0.0  
4:00 101.4 F 0.5  
6:00 138.8 F 2.3  
8:00 160.1 F 5.4  
10:00 195.8 F 8.7  
12:00 199.2 F 12.4  
14:00 213.4 F 5.0  
16:00 223.4 F 8.7  
18:00 233.8 F 13.0  
20:00 235.2 F 9.6  
22:00 238.1 F 9.9  
24:00 246.1 F 13.8  
26:00 257.4 F 19.9  
28:00 267.9 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.1 F 29.7  
2:00 272.3 F 29.3  
3:00 271.5 F 29.2  
4:00 272.7 F 29.7

Min 271.4 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.1 F 0.0  
5:00 216.1 F 0.0  
10:00 201.3 F 0.0  
15:00 194.7 F 0.0  
20:00 191.1 F 0.0  
25:00 187.6 F 0.0  
30:00 185.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:13  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:07:42  
CYCLE COMPLETE

01/16/2024 03:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_