
Midmark M11 - v1.0.5
Total Cycles: 1150

Name: RITTER2
Sterilizer ID: V2313228
06/01/2022 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.0 F 0.0
2:00 103.6 F 0.0
4:00 114.7 F 1.2
6:00 148.2 F 4.2
8:00 175.1 F 8.0
10:00 200.4 F 12.3
12:00 213.3 F 5.8
14:00 226.2 F 10.2
16:00 232.4 F 8.6
18:00 239.0 F 11.8
20:00 237.6 F 9.9
22:00 250.4 F 16.1
24:00 261.4 F 22.4
26:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.7
2:00 271.3 F 29.2
3:00 272.4 F 29.8
4:00 271.4 F 29.1

Min 271.3 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 207.7 F 0.0
10:00 215.6 F 0.0
15:00 216.9 F 0.0
20:00 216.4 F 0.0
25:00 215.2 F 0.0
30:00 212.9 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:04:32
CYCLE COMPLETE

06/01/2022 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____