
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000613
13:43 09/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:41
STERILIZING 3:45
136.0 C 220kPa 3:45
136.1 C 219kPa 4:15
136.4 C 224kPa 4:45
136.4 C 223kPa 5:15
136.3 C 224kPa 5:45
136.4 C 225kPa 6:15
136.4 C 225kPa 6:45
136.5 C 225kPa 7:15
136.5 C 225kPa 7:15
VENTING 7:15
AIR DRYING 8:08
CYCLE COMPLETE 9:19

Digital Signature #
B5AE66150E2BB27C
