
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004710
16:14 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:20
STERILIZING 3:17
136.4 C 220kPa 3:17
136.4 C 224kPa 3:47
136.4 C 224kPa 4:17
136.4 C 225kPa 4:47
136.4 C 225kPa 5:17
136.2 C 224kPa 5:47
136.4 C 225kPa 6:17
136.2 C 221kPa 6:47
136.2 C 221kPa 6:47
VENTING 6:47
AIR DRYING 7:33
CYCLE COMPLETE 14:04

Digital Signature #
421CCD69DE26BD94
