

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006936  
9:30 13/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:05  
STERILIZING 4:07  
136.4 C 222kPa 4:07  
136.4 C 222kPa 4:37  
136.4 C 223kPa 5:07  
136.4 C 224kPa 5:37  
136.4 C 224kPa 6:07  
136.4 C 225kPa 6:37  
136.4 C 224kPa 7:07  
136.0 C 222kPa 7:37  
136.0 C 222kPa 7:37  
VENTING 7:37  
AIR DRYING 8:32  
CYCLE COMPLETE 16:37

-----  
Digital Signature #  
323C12EE8262B54B  
-----