

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2671

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2022 05:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.8 F 0.0  
2:00 185.0 F 3.9  
4:00 201.0 F 9.0  
6:00 221.2 F 7.5  
8:00 236.2 F 11.1  
10:00 235.1 F 8.5  
12:00 240.8 F 9.6  
14:00 248.8 F 14.5  
16:00 261.9 F 21.9  
18:00 272.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 29.0  
2:00 271.9 F 28.3  
3:00 271.5 F 28.6  
4:00 272.7 F 29.0  
  
Min 271.2 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 238.3 F 0.0  
10:00 227.7 F 0.0  
15:00 227.7 F 0.0  
20:00 228.1 F 0.0  
25:00 228.8 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 18:17  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:49  
CYCLE COMPLETE

12/13/2022 06:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_