
Midmark M11 - v1.0.5
Total Cycles: 3546

Name: RITTER2
Sterilizer ID: V2313228
04/22/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.8 F 0.0
2:00 91.4 F 0.8
4:00 114.2 F 2.9
6:00 149.6 F 5.9
8:00 175.8 F 9.0
10:00 209.0 F 12.4
12:00 215.4 F 5.4
14:00 225.5 F 9.2
16:00 233.5 F 13.3
18:00 231.3 F 9.8
20:00 235.7 F 9.5
22:00 244.0 F 13.1
24:00 254.3 F 18.8
26:00 264.6 F 25.1
28:00 270.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 271.8 F 30.2
2:00 273.0 F 30.6
3:00 272.0 F 30.1
4:00 271.2 F 29.6

Min 271.1 F 29.6
Max 273.4 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 241.6 F 0.0
10:00 220.1 F 0.0
15:00 215.9 F 0.0
20:00 212.7 F 0.0
25:00 233.3 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 28:43
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:06:51
CYCLE COMPLETE

04/22/2024 04:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____