
Midmark M11 - v1.0.0
Total Cycles: 4284

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2024 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 175.3 F 2.2
4:00 177.3 F 5.6
6:00 192.8 F 9.7
8:00 216.5 F 8.2
10:00 219.3 F 7.1
12:00 232.8 F 11.8
14:00 233.0 F 8.5
16:00 241.2 F 11.3
18:00 244.6 F 12.6
20:00 256.4 F 18.8
22:00 266.2 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.4 F 29.3
2:00 272.3 F 29.0
3:00 271.8 F 29.0
4:00 272.5 F 29.2

Min 271.5 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 235.2 F 0.0
10:00 219.1 F 0.0
15:00 210.1 F 0.0
20:00 212.3 F 0.0
25:00 215.7 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:02:31
CYCLE COMPLETE

02/08/2024 11:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____