
Midmark M11 - v1.0.5
Total Cycles: 3407

Name: RITTER2
Sterilizer ID: V2313228
02/23/2024 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.9 F 0.0
2:00 137.7 F 0.8
4:00 153.1 F 3.4
6:00 181.5 F 7.2
8:00 200.8 F 11.3
10:00 223.5 F 5.4
12:00 228.0 F 10.1
14:00 233.1 F 9.0
16:00 240.8 F 12.0
18:00 238.9 F 10.5
20:00 251.6 F 16.6
22:00 262.5 F 23.1
24:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 272.7 F 29.8
2:00 272.2 F 30.2
3:00 271.4 F 29.6
4:00 272.6 F 30.2

Min 271.0 F 29.5
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 235.9 F 0.0
10:00 217.3 F 0.0
15:00 213.8 F 0.0
20:00 180.8 F 0.0
25:00 177.6 F 0.0
30:00 176.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:50
CYCLE COMPLETE

02/23/2024 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____