
Midmark M11 - v1.0.0
Total Cycles: 3303

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2023 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.6 F 0.0
2:00 162.4 F 0.0
4:00 185.7 F 1.8
6:00 201.0 F 6.8
8:00 218.2 F 11.8
10:00 225.6 F 7.1
12:00 238.6 F 12.6
14:00 238.7 F 10.3
16:00 235.6 F 8.3
18:00 250.7 F 15.4
20:00 262.6 F 22.2
22:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.6 F 28.1
3:00 272.5 F 29.0
4:00 272.2 F 28.6

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 211.0 F 0.0
10:00 226.9 F 0.0
15:00 231.8 F 0.0
20:00 227.8 F 0.0
25:00 223.2 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

05/30/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____