
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001653
8:24 31/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:19
PRESSURIZING 4:19
STERILIZING 5:29
136.4 C 221kPa 5:29
136.2 C 222kPa 5:59
136.4 C 220kPa 6:29
136.4 C 222kPa 6:59
136.2 C 223kPa 7:29
136.4 C 223kPa 7:59
136.0 C 221kPa 8:29
136.1 C 221kPa 8:59
136.1 C 221kPa 8:59
VENTING 8:59
AIR DRYING 9:53
CYCLE COMPLETE 9:54

Digital Signature #
EB510CC0E4859E6A
