
Midmark M11 - v1.0.0
Total Cycles: 2946

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2023 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 187.2 F 3.9
4:00 201.5 F 8.5
6:00 220.9 F 9.5
8:00 232.4 F 9.5
10:00 238.6 F 9.0
12:00 242.9 F 12.7
14:00 241.9 F 11.3
16:00 255.1 F 18.1
18:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.1
2:00 271.4 F 28.3
3:00 272.3 F 29.2
4:00 272.0 F 28.8

Min 271.1 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 221.7 F 0.0
10:00 231.8 F 0.0
15:00 225.3 F 0.0
20:00 222.5 F 0.0
25:00 222.9 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

02/27/2023 12:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____