
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005952
8:23 12/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 3:10
136.4 C 220kPa 3:10
136.4 C 224kPa 3:40
136.2 C 221kPa 4:10
136.4 C 224kPa 4:40
136.4 C 223kPa 5:10
136.4 C 223kPa 5:40
136.4 C 224kPa 6:10
136.4 C 224kPa 6:40
136.4 C 224kPa 6:40
VENTING 6:40
AIR DRYING 7:29
CYCLE COMPLETE 17:55

Digital Signature #
E39BB1E05D444C84
