
Midmark M11 - v1.0.5
Total Cycles: 292

Name: RITTER2
Sterilizer ID: V2313228
08/19/2021 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.0 F 0.0
2:00 184.5 F 3.9
4:00 205.6 F 9.4
6:00 223.3 F 6.8
8:00 235.5 F 11.8
10:00 238.2 F 10.4
12:00 238.3 F 10.0
14:00 253.5 F 17.5
16:00 266.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.0 F 28.9
2:00 272.3 F 29.3
3:00 272.1 F 29.6
4:00 271.6 F 28.7

Min 271.4 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 196.8 F 0.0
10:00 201.2 F 0.0
15:00 200.5 F 0.0
20:00 197.2 F 0.0
25:00 192.9 F 0.0
30:00 189.1 F 0.0
Drying Complete

FILLING: 3:34
HEATING: 17:28
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

08/19/2021 01:44 PM
ACCEPTED: X
REJECTED:
Operator: U02