
Midmark M11 - v1.0.0
Total Cycles: 2665

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2022 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 143.4 F 0.0
4:00 149.4 F 1.5
6:00 179.4 F 5.3
8:00 207.4 F 10.4
10:00 220.5 F 5.3
12:00 233.9 F 10.1
14:00 236.1 F 8.0
16:00 243.5 F 12.7
18:00 241.9 F 11.3
20:00 255.3 F 18.1
22:00 267.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.8
2:00 271.7 F 28.7
3:00 272.2 F 28.7
4:00 271.8 F 28.8

Min 271.4 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 231.7 F 0.0
10:00 227.9 F 0.0
15:00 228.0 F 0.0
20:00 230.5 F 0.0
25:00 231.1 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

12/13/2022 10:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____