
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001696
16:38 05/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:21
STERILIZING 3:26
136.4 C 224kPa 3:26
136.2 C 224kPa 3:56
136.2 C 225kPa 4:26
136.4 C 226kPa 4:56
136.4 C 225kPa 5:26
136.4 C 223kPa 5:56
136.0 C 222kPa 6:26
136.4 C 226kPa 6:56
136.4 C 226kPa 6:56
VENTING 6:56
AIR DRYING 7:47
CYCLE COMPLETE 7:51

Digital Signature #
33AC680D4FC865B2
