
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.4uS / 1.5ppm
CYCLE NUMBER 003740
14:11 31/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:27
STERILIZING 4:24
136.0 C 220kPa 4:24
136.4 C 224kPa 4:54
136.4 C 224kPa 5:24
136.4 C 223kPa 5:54
136.2 C 223kPa 6:24
136.4 C 223kPa 6:54
136.4 C 225kPa 7:24
136.4 C 224kPa 7:54
136.4 C 224kPa 7:54
VENTING 7:54
AIR DRYING 8:45
CYCLE COMPLETE 9:25

Digital Signature #
EDB759F4C8DD7A5D
