
Midmark M11 - v1.0.5
Total Cycles: 3945

Name: RITTER2
Sterilizer ID: V2313228
09/13/2024 01:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.4 F 0.0
2:00 118.3 F 0.0
4:00 142.1 F 0.2
6:00 175.5 F 2.1
8:00 192.1 F 5.4
10:00 205.9 F 9.0
12:00 218.8 F 12.8
14:00 225.0 F 7.3
16:00 235.8 F 11.9
18:00 235.2 F 9.2
20:00 240.3 F 10.5
22:00 246.8 F 13.8
24:00 257.9 F 20.0
26:00 268.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.6
2:00 271.9 F 28.8
3:00 271.9 F 29.3
4:00 272.7 F 29.4

Min 271.4 F 28.6
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 218.8 F 0.0
10:00 218.6 F 0.0
15:00 212.7 F 0.0
20:00 206.8 F 0.0
25:00 213.7 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 27:21
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:05:57
CYCLE COMPLETE

09/13/2024 02:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____