
Midmark M11 - v1.0.5
Total Cycles: 1762

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.1 F 0.0
2:00 170.5 F 1.5
4:00 188.7 F 6.4
6:00 209.6 F 12.0
8:00 224.2 F 7.6
10:00 237.8 F 13.0
12:00 240.8 F 11.9
14:00 241.1 F 11.3
16:00 255.3 F 18.6
18:00 267.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.6
2:00 271.7 F 28.9
3:00 272.0 F 29.5
4:00 272.4 F 29.5

Min 271.3 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 203.6 F 0.0
10:00 217.8 F 0.0
15:00 220.5 F 0.0
20:00 223.3 F 0.0
25:00 224.7 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

11/22/2022 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____