
Midmark M11 - v1.0.5
Total Cycles: 3356

Name: RITTER2
Sterilizer ID: V2313228
02/05/2024 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.6 F 0.0
2:00 127.1 F 0.2
4:00 161.2 F 2.0
6:00 174.2 F 5.2
8:00 175.0 F 9.1
10:00 193.7 F 13.2
12:00 215.7 F 6.4
14:00 228.9 F 10.7
16:00 230.5 F 7.5
18:00 239.4 F 12.5
20:00 239.3 F 11.0
22:00 251.5 F 17.0
24:00 262.4 F 23.4
26:00 272.7 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.0 F 30.2
2:00 272.5 F 29.9
3:00 272.6 F 30.5
4:00 272.0 F 29.8

Min 271.0 F 29.2
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 223.8 F 0.0
10:00 238.9 F 0.0
15:00 232.2 F 0.0
20:00 224.3 F 0.0
25:00 227.1 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:28
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:05:02
CYCLE COMPLETE

02/05/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____