

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3146

Name: RITTER2  
Sterilizer ID: V2313228  
11/20/2023 02:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.0 F 0.0  
2:00 145.4 F 1.8  
4:00 167.7 F 5.3  
6:00 190.0 F 10.0  
8:00 213.9 F 6.5  
10:00 224.7 F 9.4  
12:00 234.7 F 9.4  
14:00 241.0 F 11.9  
16:00 239.7 F 10.5  
18:00 253.6 F 17.4  
20:00 265.5 F 24.6  
22:00 272.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 271.9 F 29.2  
2:00 272.5 F 29.4  
3:00 272.7 F 29.7  
4:00 272.6 F 29.9

Min 271.5 F 29.2  
Max 273.3 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
5:00 249.4 F 0.0  
10:00 190.5 F 0.0  
15:00 186.8 F 0.0  
20:00 181.6 F 0.0  
25:00 179.3 F 0.0  
30:00 177.9 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 22:20  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:00:50  
CYCLE COMPLETE

11/20/2023 03:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_