

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004736  
13:00 21/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:05  
STERILIZING 3:07  
136.4 C 220kPa 3:07  
136.4 C 225kPa 3:37  
136.1 C 219kPa 4:07  
136.4 C 223kPa 4:37  
136.0 C 221kPa 5:07  
136.4 C 223kPa 5:37  
136.4 C 225kPa 6:07  
136.3 C 223kPa 6:37  
136.3 C 223kPa 6:37  
VENTING 6:37  
AIR DRYING 7:34  
CYCLE COMPLETE 7:47

-----  
Digital Signature #  
4B3634C3955480FA  
-----