
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005938
7:29 11/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:04
STERILIZING 4:08
136.4 C 221kPa 4:08
136.4 C 224kPa 4:38
135.9 C 219kPa 5:08
136.4 C 219kPa 5:38
136.4 C 222kPa 6:08
136.4 C 222kPa 6:38
136.4 C 222kPa 7:08
136.4 C 219kPa 7:38
136.4 C 219kPa 7:38
VENTING 7:38
AIR DRYING 8:30
CYCLE COMPLETE 9:45

Digital Signature #
167B83129B049784
