
Midmark M11 - v1.0.0
Total Cycles: 1406

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.8 F 0.0
2:00 163.5 F 1.9
4:00 178.9 F 5.8
6:00 202.7 F 9.7
8:00 227.1 F 5.2
10:00 228.6 F 8.9
12:00 240.3 F 11.7
14:00 240.5 F 10.7
16:00 235.2 F 8.3
18:00 249.7 F 14.8
20:00 260.9 F 21.1
22:00 271.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.9
2:00 271.3 F 28.3
3:00 272.7 F 29.0
4:00 271.6 F 28.3

Min 271.3 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 222.0 F 0.0
10:00 208.2 F 0.0
15:00 189.9 F 0.0
20:00 184.3 F 0.0
25:00 182.1 F 0.0
30:00 179.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 23:32
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

11/30/2021 11:41 AM
ACCEPTED: X
REJECTED:
Operator: U02