
Midmark M11 - v1.0.5
Total Cycles: 642

Name: RITTER2
Sterilizer ID: V2313228
12/16/2021 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 177.7 F 3.2
4:00 201.6 F 8.7
6:00 216.8 F 8.2
8:00 233.1 F 10.2
10:00 232.3 F 8.1
12:00 241.0 F 11.1
14:00 247.1 F 14.2
16:00 259.7 F 21.4
18:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.0
1:00 272.7 F 29.6
2:00 272.1 F 29.6
3:00 272.0 F 29.0
4:00 272.6 F 29.8

Min 271.6 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 215.3 F 0.0
10:00 240.6 F 0.0
15:00 236.9 F 0.0
20:00 226.5 F 0.0
25:00 226.2 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

12/16/2021 11:12 AM
ACCEPTED: X
REJECTED:
Operator: U02