

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4030

Name: RITTER 1  
Sterilizer ID: V2178198  
11/16/2023 01:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.5 F 0.0  
2:00 127.7 F 1.5  
4:00 148.7 F 4.5  
6:00 217.6 F 8.5  
8:00 205.9 F 13.3  
10:00 220.2 F 6.0  
12:00 229.6 F 10.5  
14:00 231.5 F 7.4  
16:00 241.4 F 12.1  
18:00 239.5 F 10.2  
20:00 252.3 F 16.3  
22:00 263.4 F 22.8  
24:00 272.3 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 271.4 F 28.5  
2:00 271.7 F 28.5  
3:00 272.4 F 28.8  
4:00 272.4 F 29.1

Min 271.3 F 28.5  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.5 F 0.0  
5:00 209.4 F 0.0  
10:00 218.2 F 0.0  
15:00 193.2 F 0.0  
20:00 188.0 F 0.0  
25:00 192.1 F 0.0  
30:00 212.3 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 24:49  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:03:23  
CYCLE COMPLETE

11/16/2023 02:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_