
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007926
8:06 07/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:45
STERILIZING 4:55
136.0 C 220kPa 4:55
136.4 C 221kPa 5:25
136.3 C 219kPa 5:55
136.4 C 218kPa 6:25
136.4 C 222kPa 6:55
136.4 C 219kPa 7:25
136.3 C 221kPa 7:55
135.9 C 220kPa 8:25
135.9 C 220kPa 8:25
VENTING 8:25
AIR DRYING 9:18
CYCLE COMPLETE 24:22

Digital Signature #
928326FB0D11DEA1
