

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3865

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2024 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.8 F 0.0  
2:00 92.1 F 0.0  
4:00 96.5 F 0.3  
6:00 122.1 F 2.6  
8:00 152.5 F 5.8  
10:00 171.6 F 9.4  
12:00 191.8 F 13.1  
14:00 213.8 F 6.1  
16:00 224.9 F 10.2  
18:00 230.5 F 9.0  
20:00 238.5 F 11.6  
22:00 237.2 F 9.7  
24:00 249.6 F 15.6  
26:00 259.9 F 21.6  
28:00 269.8 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.9  
1:00 272.2 F 29.8  
2:00 272.6 F 29.8  
3:00 272.7 F 30.0  
4:00 272.8 F 30.2  
  
Min 271.9 F 29.8  
Max 273.2 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
5:00 215.3 F 0.0  
10:00 207.2 F 0.0  
15:00 188.8 F 0.0  
20:00 183.6 F 0.0  
25:00 183.2 F 0.0  
30:00 195.2 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 29:45  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:08:22  
CYCLE COMPLETE

08/15/2024 03:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_