

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002557  
10:30 04/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:50  
STERILIZING 3:59  
136.4 C 224kPa 3:59  
136.2 C 221kPa 4:29  
136.3 C 222kPa 4:59  
136.4 C 224kPa 5:29  
136.3 C 224kPa 5:59  
136.4 C 225kPa 6:29  
136.4 C 223kPa 6:59  
136.4 C 225kPa 7:29  
136.4 C 225kPa 7:29  
VENTING 7:29  
AIR DRYING 8:24  
CYCLE COMPLETE 8:31  
-----

Digital Signature #  
3FF2E7744D84ACD6  
-----