
Midmark M11 - v1.0.0
Total Cycles: 5436

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2025 01:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.0 F 0.0
2:00 135.7 F 2.1
4:00 144.0 F 5.0
6:00 168.1 F 8.5
8:00 202.8 F 13.0
10:00 211.0 F 5.1
12:00 223.3 F 9.3
14:00 238.8 F 11.5
16:00 238.0 F 10.2
18:00 236.9 F 8.7
20:00 247.7 F 14.1
22:00 258.8 F 20.3
24:00 269.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.0 F 29.3
2:00 272.3 F 29.0
3:00 272.3 F 29.6
4:00 272.2 F 29.1

Min 270.9 F 28.3
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 208.4 F 0.0
10:00 198.4 F 0.0
15:00 214.8 F 0.0
20:00 222.5 F 0.0
25:00 222.9 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:57
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

03/24/2025 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____