

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002070  
10:21 20/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:44  
PRESSURIZING 1:51  
STERILIZING 2:48  
136.0 C 220kPa 2:48  
136.4 C 225kPa 3:18  
136.0 C 221kPa 3:48  
136.4 C 227kPa 4:18  
136.4 C 227kPa 4:48  
136.2 C 224kPa 5:18  
136.3 C 225kPa 5:48  
136.0 C 224kPa 6:18  
136.0 C 224kPa 6:18  
VENTING 6:18  
AIR DRYING 7:06  
CYCLE COMPLETE 37:07  
-----

Digital Signature #  
051F7423182C9E1C  
-----