
Midmark M11 - v1.0.0
Total Cycles: 4705

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2024 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.2 F 0.0
2:00 138.4 F 1.9
4:00 154.9 F 5.2
6:00 180.9 F 9.2
8:00 195.5 F 11.7
10:00 211.9 F 6.1
12:00 225.2 F 10.6
14:00 228.0 F 7.0
16:00 239.4 F 11.7
18:00 237.0 F 9.2
20:00 250.3 F 15.5
22:00 261.6 F 21.9
24:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.5 F 29.3
4:00 272.0 F 28.7

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 204.7 F 0.0
10:00 198.4 F 0.0
15:00 206.8 F 0.0
20:00 223.5 F 0.0
25:00 229.6 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:51
CYCLE COMPLETE

07/08/2024 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____